



27 May 2022

This memo is to notify you regarding the extension of the shelf-life expiration date of the INDICAID COVID-19 Rapid Antigen Test. As a part of the EUA conditions of authorization, PHASE Scientific submits updated real-time product stability data to the U.S. Food and Drug Administration (FDA) on an on-going basis.

Pursuant to FDA's Policy for Coronavirus Disease-2019 Tests During the Public Health Emergency (Revised), Guidance for Developers and Food and Drug Administration Staff (November 15, 2021 revision), we have updated the shelf-life expiration to twelve (12) months at 2-30°C from the date of manufacture. Please see FDA letter RE: EUA210259/S004.

The shelf-life for all lots of INDICAID with initial prefix 21SXXXX and lots 22S0001-22S0100 may be extended by six (6) months from the expiration dating printed on the product packaging.

Felix Chao  
VP, US General Manager

*Felix Chao*

---